

Analgesy-Meter

Randall-Selitto Paw Pressure Test

Cat. No. 37215

General

The 37215 is the up to date version of the classical 7200 paw pressure test which, **since 1965**, is helping to perform a rapid precise screening of analgesic drugs in a number of academic and industrial laboratories.

The force is applied to the animal's paw, which is placed on a small plinth under a cone-shaped pusher with a rounded tip.

The operator depresses a pedal switch to start the mechanism which exerts the force.

When the rat struggles, the operator releases the pedal and reads off the scale the force at which the animal felt pain.

NEW: we are now introducing a **specific pressure sensor and the related controller, available as optional, to transform the Analgesy-Meter in a fully digital device.**

As the basic design is unchanged, results with the digital model are **consistent with published data.**

The upgrade kit has been designed to be fitted on existing Ugo Basile Analgesy-Meters as well. Ask for details!

NEW



now available with
optional upgrade to
digital reading

Main Features

- Same instrument, three force ranges (from 0 to 250, 500, 750 g)
- Simple and reliable: no calibration needed!
- **NEW model with digital reading**
- Specific version for Mouse available, with lower (50% pressure range)
- Classic method since the 1960s: hundreds of papers published!
- **Upgrade kit for old Analgesy-Meters available**

Instrument Description

The force applied to the paw by the plinth increases at a constant rate, thus enabling perfect reproducible measurements to be made. The motor stops immediately the pedal is released.

The force is measured on the scale calibrated in 10-gram steps, by a pointer riveted to the slide. The scale can be multiplied by 2 or 3, by placing on the slide one or two discs provided with the standard package.

After each test the slide should be returned to its starting point by lifting it and pushing it to the left.

The 37215 features a low voltage synchronous motor and conforms the CE rules.

The standard 37215 can be conveniently used with mice. However, a dedicated model is also available, when lower pressure (50%) is desirable, model **37216**, which includes a special chisel-shaped pusher (also available separately)

Data Acquisition

The classic Analgesy-Meter can now be integrated with a **specific pressure sensor and the related controller, available as optional, which upgrades the Analgesy-Meter to a fully digital device.**



As the basic design is unchanged, results with the digital model are **consistent with published data.**

The design of the upgrade kit makes it easy to retrofit existing UB Analgesy-Meters as well.

Ask for details!

37215 Specifications

Power Requirements: 115 or 230V, 50/60Hz, 15W max.

Start / Stop : by pedal switch

Force Range 37215 : 0 to 250, 500, 750 grams

37216 : 0 to 125, 250, 375 grams

Physical:

Dimensions : cm 40 x 16 x 14

Packing : cm 55 x 45 x 36

Weight : 2.1Kg

Shipping Weight : 5.0Kg approx

Ordering Information

37215 **ANALGESY-METER**, complete with following standard accessories:-

37215-302 Instruction Manual (on USB key)

37215-303 Pedal Switch, complete with cable

37215-323 Set of discs for additional weight

37215-321 Plinth

37215-322 Standard Pusher *

E-WP008 Mains Cord

* Pushers in special material/shapes, available on request

37216 **ANALGESY-METER**, low-pressure model, suitable for mice, with pusher 37215-326

Optional Upgrade to Digital

37215-100 **ANALGESY DAQ** upgrade kit

37215-BUNDLE Analgesy-Meter & Upgrade Kit

Bibliography

METHOD PAPER

- L.O. Randall and J.J. Selitto: "A Method for Measurement of Analgesic Activity on Inflamed Tissue" *Arch. Int. Pharmacodyn. CXI*, No. 4: 409-419, 1957.

REFERENCE TO UB ANALGESY-METER (RAT)

- E.K. Joseph et alia: "Vascular Endothelial Cells Mediate Mechanical Stimulation-Induced Enhancement of Endothelin Hyperalgesia via Activation of P2X2/3 Receptors on Nociceptors" *J. Neuroscience* 33 (7): 2849-2859, 2013
- L. Ferrari et alia: "Role of Nociceptor α CaMKII in Transition from Acute to Chronic Pain (Hyperalgesic Priming) in Male and Female Rats" *J. Neuro-science* 33 (27): 11002-11011, 2013
- D.A. Andersson et alia: "TRPA1 Has a Key Role in the Somatic Pro-Nociceptive Actions of Hydrogen Sulfide" *PLoS ONE* 7(10): e46917, 2012
- K. Király et alia: "The Dipeptidyl Peptidase IV (CD26, EC 3.4.14.5) Inhibitor Vildagliptin is a Potent Antihyperalgesic in Rats by Promoting Endomorphin-2 Generation in the Spinal Cord" *Eur. J. Pharmacol.* 650: 195-199, 2011
- Zs. Helyes et alia: "Involvement of Transient Receptor Potential Vanilloid 1 Receptors in Protease-Activated Receptor-2-induced Joint Inflammation and Nociception" *Eur. J. of Pain* 14 (4): 351-358, 2010

REFERENCE TO UB ANALGESY-METER (MOUSE)

- K. Sugimoto et alia: "The Impact of Low-Dose Insulin on Peripheral Nerve Insulin Receptor Signaling in Streptozotocin-Induced Diabetic Rats" *PLoS ONE*: 8(8): e74247, 2013
- M.J. Hussey et alia: "Deletion of the Adenosine A2A Receptor in Mice enhances Spinal Cord Neurochemical Responses to an Inflammatory Nociceptive Stimulus" *Neuroscience Letters* 506(2): 198-202, 2012
- M.S. Nash et alia: "7-tert-Butyl-6-(4-Chloro-Phenyl)-2-Thioxo-2,3-Dihydro-1H-Pyrido[2,3-d]Pyrimidin-4-One, a Classic Polymodal Inhibitor of ..." *J. Pharmacol. Exper. Therap.* 342 (2): 389-398, 2012